



Salvi Casagrande
Medição e Automação. Qualidade Absoluta.

REFRATÔMETRO



TIPO	MODELO	ESCALA	RESOLUÇÃO	MEDIDAS (mm)
Brix	REF 101	0~10%	0.1%	27 x 40 x 210
	REF 102	0~18%	0.1%	27 x 40 x 190
	REF 103	0~32%	0.2%	27 x 40 x 160
	REF 104	28~62%	0.2%	27 x 40 x 148
	REF 105	45~82%	0.5%	27 x 40 x 140
	REF 106	58~90% Brix	0.5% Brix	27 x 40 x 160
		38~43° Be'	0.5° Be'	
		17~27% Água	1% Água	
	REF 107	0~90% (3 escalas)	0.2%	27 x 40 x 200
	REF108	0~80% (2 escalas)	0.1%	ø 31 x 175
Brix/ ATC (Compensação Automática da Temperatura)	REF 111	0~10%	0.1%	29 x 40 x 210
	REF 113	0~32%	0.2%	29 x 40 x 160
	REF 114	28~62%	0.2%	29 x 40 x 148
	REF 116	58~90% Brix	0.5% Brix	29 x 40 x 160
		38~43° Be'	0.5° Be'	
Salinidade	REF 201	0~100%	1%	27 x 40 x 190
	REF 202	0~28%	0.2%	27 x 40 x 160
Salinidade/ ATC	REF 211	0~100%	1%	29 x 40 x 190
	REF 212	0~28%	0.2%	29 x 40 x 190
Proteína	REF 301	0~12 g/dl	0.2 g/dl	27 x 40 x 160
		1000~1040 sg	0.002 sg	
	REF 302	0~12 g/dl	0.2 g/dl	27 x 40 x 160
		1000~1050 sg	0.002 sg	
Proteína/ ATC	REF 312	1.3330~1.3600 RI	0.00025 ou 0.0005 RI	27 x 40 x 160
		1.3330~1.3600 RI	0.00025 ou 0.0005 RI	
	REF 401	-60~30°F	10°F	27 x 40 x 160
		1.15~1.30 sg	0.01 sg	
Bateria/ Resfriamento	REF 402	-50°C~0°C	10°C	27 x 40 x 160
		1.15~1.30 sg	0.01 sg	
	REF 403	-50°C~0°C	10°C	27 x 40 x 160
		-40°C~0°C	5°C	
Alcool	REF 501	0~80% w/w	1.0%	27 x 40 x 160
	REF 502	0~25% vol	0.2%	29 x 40 x 160
Alcool/ ATC	REF 512	1.30~1.81	1.0%	35 x 70 x 130
Gemological	REF 601	1.30~1.81	1.0%	35 x 70 x 130
Brix/ ATC	REF 702	0~32%	1.0%	29 x 40 x 160
Brix/ ATC	REF 712	0~32%	1.0%	29 x 40 x 160



Salvi Casagrande
Medição e Automação. Qualidade Absoluta.



(11) 3327-0033



+55 11 3327-0035



vendas@salvicasagrande.com.br



www.salvicasagrande.com.br



@salvi.casagrande



R. Florêncio de Abreu, 605 - Centro - São Paulo, SP.

